



IN APPLIED SCIENCE & ENGINEERING TECHNOLOGY

Volume: 10 Issue: V Month of publication: May 2022

DOI: https://doi.org/10.22214/ijraset.2022.42799

www.ijraset.com

Call: 🛇 08813907089 🕴 E-mail ID: ijraset@gmail.com



Design of Mattress Compatible for MRI/CT Scans

Manasi Wagh¹, Rutuja Pawar², Prateek Kohinkar³, Harshal Koli⁴

^{1, 2, 3, 4}Student, Department of Mechanical Engineering, JSPM's Rajarshi Shahu college of engineering, Tathawade, Maharashtra, India.

I. INTRODUCTION

Humans normally maintain a body temperature of 37°C, but we know in specific scans like CT/MRI/PET scan exam rooms the temperature needs to be maintained at around 18°C to 22°C. Here the body temperature and room temperature don't match this leads to the negative or minor incorrect outcome. There-fore we have to design a mattress which is fully compatible for MRI scans. This will allow us to set the mattress temperature in between the range of 20°C to 40°C. It will mainly consist of devices like fluid pump, temperature sensors, multipurpose heat pump, micro controller and many other such devices.

II. METHODOLOGY

The mattress has dimensions as (2×0.42) meter. It has two openings: one for the inlet of water and the other one for the outlet of water. The mattress is designed in such a way that it allows the water to flow through it and eventually making the mattress warmer on the outside that is the surface, where the patient's body will rest. The temperature of the mattress can be adjusted according to the user or the requirements of the staff in between the range of 20°C to 40°C with the help of provided input keypad. TS1 and TS2 are the temperature sensors which are placed at the inlet and the outlet of the mattress respectively to sense the temperature of the fluid. The mattress is designed with the smart mechanism that it automatically heats the water when the once set temperature by the user falls down. The heat pump, temperature sensors, relay, cooling pump and input keypad are interfaced with the microcontroller.



III. MODELING AND ANALYSIS



International Journal for Research in Applied Science & Engineering Technology (IJRASET)

ISSN: 2321-9653; IC Value: 45.98; SJ Impact Factor: 7.538 Volume 10 Issue V May 2022- Available at www.ijraset.com

Technology	Stack	
Micro controller	Node MCU/Node32/Arduino	
Plastic tube [for inlet and outlet]	SY CE RoHs plastic pipe	
Fluid pump	Stainless steel fluid pump 140 LPM	Node MCU
Multipurpose heat pump	AVI-WHH-25LDX Horizontal heat pump	
Temperature sensors	DS18B20 sensor	G
LCD Display	Locally available	
Temperature input key pad	4*4 matrix keypad	Temp. Sensor
Matress material	Rubber	
relay	Local relay	





Fluid pump

RESULTS AND DISCUSSION IV.

By using the mentioned components we have achieved the final results. User needs to set the temperature according to their convenience. Microcontroller receives the input and the relay is ON. Heat pump heats the water from the water reservoir. This heated water then enters the mattress. Temperature sensor then senses the temperature at the outlet. If the temperature at the outlet is Ts2<Ts1 then Relay and pump is set ON.

V. CONCLUSION

The main objective of this research was to design and develop rubber mattress for MRI/CT/PET scans. The design and material of the mattress has been approved by a MRI technician. The temperature can also be controlled with the help of mobile phone application or laptop by connecting the microcontroller with Wi-fi. Sensors, fluid pump, heat pump used for the purpose are MRI compatible.

REFERENCES

- [1] https://en.wikipedia.org/wiki/NodeMCU
- https://elearn.ellak.gr/mod/book/view.php?id=2326 [2]
- https://fabotics.in/product/nodemcu-esp8266-wifi-development-board/ [3]
- https://pe-ltd.com/product/engine-coolant-temperature-sensor/ [4]
- [5] https://ph.parker.com/us/en/diaphragm-pumps











45.98



IMPACT FACTOR: 7.129







INTERNATIONAL JOURNAL FOR RESEARCH

IN APPLIED SCIENCE & ENGINEERING TECHNOLOGY

Call : 08813907089 🕓 (24*7 Support on Whatsapp)