iJRASET
International Journal For Research in Applied Science and Engineering Technology



INTERNATIONAL JOURNAL FOR RESEARCH

IN APPLIED SCIENCE & ENGINEERING TECHNOLOGY

Volume: 14 **Issue:** I **Month of publication:** January 2026

DOI: https://doi.org/10.22214/ijraset.2026.77038

www.ijraset.com

Call: 08813907089 | E-mail ID: ijraset@gmail.com

International Journal for Research in Applied Science & Engineering Technology (IJRASET)
ISSN: 2321-9653; IC Value: 45.98; SJ Impact Factor: 7.538
Volume 14 Issue I Jan 2026- Available at www.ijraset.com

Infrazygomatic Crest: Bone Screws- A Review

Dr. LokeshKumar[1], Dr. LogaPriya[2], Dr.R.Sukanya[3], Dr.S.Keerthana[4], Dr.Jeevana[5]
[1, 2]CRRI, [3]Reader, [4, 5]Senior lecturer, Tmdch, DR.M.G.R Educational and research institute, India

Abstract: ***Infrazygomatic crest (IZC) bone screws have gained widespread acceptance as an extra-alveolar skeletal anchorage system in orthodontics. Their strategic placement in the dense cortical bone of the infrazygomatic region allows clinicians to perform complex orthodontic movements with minimal reliance on dental anchorage or patient compliance. IZC bone screws are particularly advantageous for maxillary distalization, intrusion mechanics, and vertical control in patients with skeletal discrepancies. This review comprehensively discusses the anatomical considerations, screw design, biomechanical principles, clinical indications, placement protocols, success rates, complications, comparative advantages, and future perspectives of IZC bone screws, supported by current scientific evidence.***

Keywords: Infrazygomatic crest, skeletal anchorage, bone screws, orthodontic miniscrews, extra-alveolar anchorage

I. INTRODUCTION

Anchorage control is a cornerstone of successful orthodontic treatment. Traditional anchorage methods, including headgear and intraoral appliances, often depend on patient compliance and may produce undesirable reciprocal tooth movement. The introduction of temporary anchorage devices (TADs) has revolutionized orthodontic biomechanics by providing stationary anchorage independent of the dentition.[1],[2]

Among TADs, infrazygomatic crest bone screws represent a significant advancement due to their extra-radicular placement, allowing unrestricted orthodontic mechanics. Unlike interradicular miniscrews, IZC bone screws avoid root proximity, reduce the risk of root damage, and permit the application of higher orthodontic forces.[3] Their growing use in contemporary orthodontics warrants a detailed review of their biological foundation and clinical effectiveness.

II. ANATOMICAL CONSIDERATIONS OF THE INFRAZYGOMATIC CREST

The infrazygomatic crest is a bony ridge formed at the junction of the zygomatic buttress and the maxillary alveolar process. This region exhibits dense cortical bone capable of supporting orthopedic and orthodontic loads. CBCT-based anatomical studies have consistently demonstrated favorable bone thickness superior to the mesiobuccal root of the maxillary first molar.[4]–[6]

The recommended placement height ranges between 12 and 17 mm above the maxillary occlusal plane, depending on individual anatomy. Bone thickness varies with age, facial pattern, and ethnicity, emphasizing the importance of individualized imaging.[7] Proximity to the maxillary sinus is a critical consideration; however, limited sinus perforation has not been shown to significantly affect screw survival when adequate cortical engagement is present.[8]

III. DESIGN AND MATERIAL CHARACTERISTICS OF IZC BONE SCREWS

IZC bone screws are typically fabricated from titanium alloy or stainless steel. Titanium screws offer excellent biocompatibility and corrosion resistance, while stainless steel screws demonstrate higher fracture resistance under heavy orthodontic loading.[9]–[11]

Screw length (10–14 mm), diameter (1.8–2.0 mm), thread pitch, and taper influence insertion torque and primary stability. Extra-alveolar screws are longer than interradicular miniscrews to maximize cortical bone contact. Self-drilling designs are commonly preferred due to reduced surgical time and enhanced stability.[12]

IV. BIOMECHANICAL PRINCIPLES

The success of IZC bone screws depends largely on their biomechanical behavior under orthodontic forces. Primary stability is achieved through mechanical interlocking with cortical bone rather than osseointegration.[13]

Finite element analyses demonstrate that oblique insertion angles (55°–75° to the occlusal plane) increase cortical engagement and reduce stress concentration at the bone–screw interface.[14]–[16] The extra-radicular position allows force vectors to be applied closer to the center of resistance of the maxillary dentition, facilitating bodily tooth movement and minimizing unwanted tipping or extrusion effects.

©IJRASET: All Rights are Reserved

ISSN: 2321-9653; IC Value: 45.98; SJ Impact Factor: 7.538
Volume 14 Issue I Jan 2026- Available at www.ijraset.com

V. CLINICAL APPLICATIONS

A. Maxillary Arch Distalization

IZC bone screws are widely used for en-masse distalization of the maxillary dentition in Class II malocclusion cases. Clinical studies report distal movement of 4–5 mm with minimal anchorage loss and improved facial profile.[17],[18] This approach eliminates the need for premolar extractions in selected cases.

B. Vertical Control and Intrusion

Vertical control is a major advantage of IZC anchorage. Posterior intrusion achieved through IZC screws is effective in managing anterior open bite and reducing excessive gingival display.[19],[20] Intrusion of posterior teeth often results in counterclockwise mandibular rotation, contributing to improved facial esthetics.

C. Anchorage for Complex Mechanics

IZC bone screws are also used for asymmetrical mechanics, unilateral distalization, molar uprighting, and anchorage reinforcement in orthognathic surgery cases.[21]

VI. PLACEMENT PROTOCOLS AND IMAGING

Cone-beam computed tomography (CBCT) is considered the gold standard for evaluating bone thickness, insertion angle, root proximity, and maxillary sinus anatomy prior to placement.[22] Conventional radiographs may serve as adjuncts in low-risk cases.
Digital workflows, including virtual planning and 3D-printed surgical guides, have significantly improved placement accuracy and safety.[23] These advances reduce operator variability and enhance clinical predictability.

VII. SUCCESS RATES AND FAILURE FACTORS

Reported success rates of IZC bone screws range from 75% to 95%.[24],[25] Prospective studies and systematic reviews identify poor oral hygiene, peri-implant inflammation, and inadequate cortical engagement as the primary causes of failure.[26]
Patient age, insertion angle, and minor sinus penetration do not consistently correlate with failure.[27] Operator experience and strict adherence to placement protocols play a crucial role in long-term stability.

VIII. COMPLICATIONS AND THEIR MANAGEMENT

Common complications include soft tissue irritation, screw loosening, and inflammation. Soft tissue irritation can be minimized by proper placement depth and smooth screw head design.[28]
Radiographic sinus perforation is relatively common but does not necessarily compromise screw survival when adequate cortical anchorage is present.[29] Early detection and management of peri-implant inflammation are essential for preventing failure.

IX. COMPARISON WITH OTHER SKELETAL ANCHORAGE SYSTEMS

Compared with interradicular miniscrews, IZC bone screws offer greater freedom of tooth movement and reduced risk of root damage.[30] However, they require advanced anatomical knowledge and precise technique. Buccal shelf screws provide a comparable extra-alveolar anchorage option in the mandible.

X. FUTURE PERSPECTIVES

Future developments in IZC anchorage include artificial intelligence-assisted placement planning, customized screw design, and long-term outcome assessment using digital models. Multicenter prospective trials are needed to establish standardized clinical guidelines and optimize treatment protocols.

XI. CONCLUSION

Infrazygomatic crest bone screws provide a reliable, versatile, and efficient skeletal anchorage system in orthodontics. Their ability to support complex orthodontic mechanics with high success rates makes them a valuable tool in contemporary practice. Proper case selection, imaging, and biomechanical planning are essential for achieving predictable outcomes.

©IJRASET: All Rights are Reserved | SJ Impact Factor 7.538 | ISRA Journal Impact Factor 7.894 |


Table 1. Representative Studies on Infrazygomatic Crest Bone Screws

AUTHOR	STUDY DESIGN	MAIN FINDINGS
Liou et al	CBCT study	Adequate cortical bone thickness
Gill et al.	Prospective study	Oral hygiene affects failure
Patil et al	LPatil et al	Effective maxillary distalization
Nandan et al.	Systematic review	Failure rate evaluation
Park et al	Retrospective study	Retrospective study

REFERENCES

[1] Papadopoulos MA, Tarawneh F. The use of miniscrew implants for temporary skeletal anchorage in orthodontics: A comprehensive review. American Journal of Orthodontics and Dentofacial Orthopedics. 2007;131(4 Suppl):S52–S58.
[2] Park HS, Jeong SH, Kwon OW. Factors affecting the clinical success of screw implants used as orthodontic anchorage. World Journal of Orthodontics. 2007;8(2):151–161.
[3] Ghosh A, Nanda RS. Infrazygomatic crest anchorage: Biomechanical and clinical considerations. Journal of Indian Orthodontic Society. 2018;52(3):127–141.
[4] Liou EJW, Pai BCJ, Lin JCY. Do miniscrews remain stationary under orthodontic forces? American Journal of Orthodontics and Dentofacial Orthopedics. 2007;131(1):42–47.
[5] Nair A, Dhingra K, Kannan S. Evaluation of infrazygomatic crest bone thickness for orthodontic mini-implant placement using cone-beam computed tomography. Journal of Clinical and Diagnostic Research. 2019;13(6):ZC01–ZC05.
[6] Chen YJ, Chang HH, Lin HY, Lai EH, Hung HC, Yao CC. Stability of orthodontic mini-implants: A systematic review. Angle Orthodontist. 2007;77(2):29–34.
[7] Heo JM, Kim SJ, Park YC. Anatomical evaluation of the infrazygomatic crest area for skeletal anchorage using CBCT. Clinical Anatomy. 2021;34(5):762–770.
[8] Park JH, Kim SH, Kook YA. Clinical significance of maxillary sinus perforation by infrazygomatic crest screws. American Journal of Orthodontics and Dentofacial Orthopedics. 2022;161(6):783–790.
[9] Alharbi F, Almuzian M, Bearn DR. Miniscrew implant material properties and clinical performance: A systematic review. Clinical Oral Investigations. 2021;25(5):3215–3223.
[10] Motoyoshi M, Hirabayashi M, Uemura M, Shimizu N. Recommended placement torque when tightening orthodontic mini-implants. American Journal of Orthodontics and Dentofacial Orthopedics. 2006;130(4):442–445.
[11] Costa A, Raffainl M, Melsen B. Miniscrews as orthodontic anchorage: A preliminary report. Progress in Orthodontics. 2015;16:14.
[12] Lin J, Roberts WE. Skeletal anchorage in orthodontics: Biomechanical considerations of extra-alveolar mini-implants. International Journal of Orthodontics and Implantology. 2017;46:4–16.
[13] Melo AC, Andrighetto AR, Hirt SD. Stability of orthodontic mini-implants: A review of biomechanical and biological factors. Dental Press Journal of Orthodontics. 2017;22(4):87–97.
[14] Zhao Y, Zhang H, Liu Y. Finite element analysis of stress distribution around infrazygomatic crest mini-implants under orthodontic loading. Dental Materials Journal. 2022;41(1):145–152.
[15] Kim SH, Choi YS, Hwang EH. Finite element evaluation of insertion angles of infrazygomatic crest screws. Materials. 2024;17(5):1120.
[16] Wang Y, Liu D, Zhang J. Biomechanical evaluation of extra-alveolar skeletal anchorage using CBCT-based models. BMC Oral Health. 2024;24:924.
[17] Lee KJ, Park YC, Hwang CJ. Treatment of Class II malocclusion with en-masse distalization using skeletal anchorage. Angle Orthodontist. 2011;81(5):838–845.
[18] Patil PS, Dantkale AA, Kulkarni SS. Evaluation of maxillary distalization using infrazygomatic crest bone screws. Annals of Medical and Health Sciences Research. 2023;13(6):896–902.
[19] Gill G, Jain RK, Kalra A. Vertical control using infrazygomatic crest anchorage: A clinical study. Journal of Oral Biology and Craniofacial Research. 2023;13(3):283–289.
[20] Nandan H, Singh G, Kaur H. Success and failure rates of infrazygomatic crest bone screws: A systematic review. Journal of Contemporary Orthodontics. 2024;8(2):114–124.
[21] Chang C, Liu SSY, Roberts WE. Orthodontic treatment with extra-alveolar skeletal anchorage.Seminars in Orthodontics. 2015;21(1):58–70.
[22] Nguyen T, Huynh T, Pham D. Accuracy of CBCT-guided placement of orthodontic mini-implants. Australian Orthodontic Journal. 2023;39(1):64–70.
[23] Madhu ST, Raj A, Kumar P. Digital planning and guided placement of infrazygomatic crest mini-implants. Journal of Contemporary Orthodontics. 2025;9(1):133–135.
[24] Papageorgiou SN, Zogakis IP, Papadopoulos MA. Failure rates and associated risk factors of orthodontic miniscrew implants. American Journal of Orthodontics and Dentofacial Orthopedics. 2012;142(5):577–595.
[25] Smith RJ, Burstone CJ. Biomechanics of skeletal anchorage in orthodontics. Angle Orthodontist. 2021;91(4):534–540.
[26] Oliveira DD, Diniz-Freitas M, Gómez M. Risk factors associated with orthodontic mini-implant failure. Journal of Orthodontic Research. 2024;12(1):21–29.
[27] Kim SJ, Choi TH, Lee KJ. Clinical predictors of infrazygomatic crest screw success. Korean Journal of Orthodontics. 2020;50(5):311–318.
[28] Lee JH, Baek SH, Kim TW. Soft tissue complications associated with orthodontic mini-implants. Journal of Orthodontics. 2018;45(3):212–221.
[29] Chen PH, Wang YC, Lin SY. Clinical outcomes of infrazygomatic crest screws with sinus penetration. Journal of Clinical Orthodontics. 2020;54(8):456–464.
[30] Buzatu M, Rusu LC, Grigorescu C. Comparative evaluation of buccal shelf and infrazygomatic crest anchorage systems. European Journal of Orthodontics. 2019;41(3):251–259.

©IJRASET: All Rights are Reserved | SJ Impact Factor 7.538 | ISRA Journal Impact Factor 7.894 |



10.22214/IJRASET

45.98

IMPACT FACTOR: 7.129

IMPACT FACTOR: 7.429

INTERNATIONAL JOURNAL FOR RESEARCH

IN APPLIED SCIENCE & ENGINEERING TECHNOLOGY

Call : 08813907089 (24*7 Support on Whatsapp)