iJRASET
International Journal For Research in Applied Science and Engineering Technology

INTERNATIONAL JOURNAL FOR RESEARCH

IN APPLIED SCIENCE & ENGINEERING TECHNOLOGY

Volume: 13 **Issue:** X **Month of publication:** October 2025

DOI: https://doi.org/10.22214/ijraset.2025.74767

www.ijraset.com

Call: 08813907089 | E-mail ID: ijraset@gmail.com

International Journal for Research in Applied Science & Engineering Technology (IJRASET)
ISSN: 2321-9653; IC Value: 45.98; SJ Impact Factor: 7.538
Volume 13 Issue X Oct 2025- Available at www.ijraset.com

Posture Correction Device Using ESP32 and MPU6050

Mrs. A. Rajitha[1], Nidhii Mallavarapu[2], P. S. V. Susmitha[3], S. Roshini[4], P. Tejaswi[5]
[1]*Assistant Professor, Department of ETM, G. Narayanamma Institute of Technology and Science, Hyderabad, India*
[2, 3, 4, 5]*Bachelor of Technology Students, Department of Electronics and Telematics Engineering, G. Narayanamma Institute of Technology and Science, Hyderabad, India*

Abstract: ***Prolonged sitting, especially among students and professionals, often leads to poor posture, leading to chronic back pain and musculoskeletal issues. This project proposes a posture correction device leveraging the MPU6050 sensor and ESP32 microcontroller. The system continuously monitors spinal orientation and alerts the user via a buzzer and OLED display when slouching is detected. Unlike traditional methods such as ergonomic furniture or physical therapy, this device offers real-time, non-intrusive feedback. Its lightweight, affordable design encourages daily use, promoting healthier posture habits and minimizing long-term health risks.***

Keywords: ESP32, MPU6050, Posture Monitoring, Wearable Device, Alert System, Buzzer, OLED.

I. INTRODUCTION

Poor posture has become prevalent due to the excessive use of digital devices and prolonged sedentary lifestyles. Traditional posture correction methods—ergonomic furniture, physiotherapy, or wearable belts—either lack interactivity or require external supervision. With advances in embedded systems and sensor technology, posture monitoring can now be implemented in real-time. The proposed system uses an MPU6050 (3-axis accelerometer and gyroscope) to monitor back alignment, and an ESP32 microcontroller to process the sensor data. The system issues alerts via a buzzer and OLED display when posture deviates beyond a threshold. Its small size makes it suitable among students, office workers, and the elderly.

II. LITERATURE SURVEY

Several research studies emphasize the relevance and benefits of wearable posture monitoring systems:
- Hansraj (2004) and Liao (2016) identify the postural health risks associated with digital lifestyle.
- Uribe-Quevedo (2013) and Cornell University projects highlight the efficacy of real-time feedback through haptic alerts.
- Singh et al. (2019) demonstrated ESP32-based implementations using MPU6050 for cost-effective, real-time monitoring.
- Park et al. (2020) validated posture wearables for educational and workplace applications.
- Kendall et al. (2005) emphasized the need for user comfort to ensure regular usage.

III. METHODOLOGY

The system architecture integrates the MPU6050 sensor with ESP32 to detect angular tilt. When slouching is detected, the system provides audio and visual alerts through a buzzer and OLED display respectively.



Fig1: ESP32

©IJRASET: All Rights are Reserved



Fig2: MPU6050



Fig3: Block diagram

IV. IMPLEMENTATION

The embedded system is programmed using the Arduino IDE. The code initializes the MPU6050, continuously reads acceleration data, calculates pitch, and triggers alerts if the posture threshold is breached. Bluetooth capability is included for potential future extensions like mobile alerts or logging in addition with visual and audio feedback.



Fig4: Prototype

©IJRASET: All Rights are Reserved | SJ Impact Factor 7.538 | ISRA Journal Impact Factor 7.894 |

V. RESULTS

The device was assembled as a compact wearable with the MPU6050 placed on the upper back. When tested, the system effectively detected poor posture and issued alerts within milliseconds. The OLED displayed appropriate messages, and the buzzer provided timely audio feedback.



Fig5: Output - 1



Fig6 : Output – 2

VI. CONCLUSION

The posture correction system presented is a lightweight, low-cost, and real-time solution for improving postural habits. With reliable tilt detection using MPU6050 and effective alerts via buzzer and OLED, the device demonstrates good potential for practical daily use.

REFERENCES

[1] Hansraj, K. (2004). “Assessment of stresses in the cervical spine.” Surgical Technology International.
[2] Liao, D.Y. (2016). “Biofeedback Head-and-Neck Posture Correction.” IEEE Conf. on Connected Health.
[3] Uribe-Quevedo, A., et al. (2013). “Seated tracking for correcting computer work postures.”
[4] Singh, A., et al. (2019). “ESP32-based Posture Correction Systems.”
[5] Park, et al. (2020). “Smart Wearables for Posture Monitoring.”
[6] Kendall et al. (2005). “Posture and Pain: An Ergonomic Evaluation.”
[7] Lee, H., Kim, J., & Park, S. (2018). “Development of a wearable posture correction system using IMU sensors.” *IEEE Sensors Journal*, 18(16), 6718–6725.
[8] Chae, S. H., & Park, J. (2017). “Real-time posture monitoring system using flex sensors and Arduino.” *Proceedings of the International Conference on Control, Automation and Systems (ICCAS)*, 1552–1555.
[9] Gupta, P., & Srivastava, S. (2020). “Smart wearable posture correction device using IoT.” *International Journal of Advanced Research in Computer and Communication Engineering (IJARCCE)*, 9(3), 45–49.
[10] Wang, X., Zhao, Y., & Lin, F. (2019). “A wearable posture monitoring system for rehabilitation using accelerometer and gyroscope.” *IEEE Transactions on Instrumentation and Measurement*, 68(8), 2753–2762.
[11] Nguyen, T. H., & Lee, D. (2021). “Posture monitoring and correction using machine learning and IMU-based wearable sensors.” *Sensors*, 21(2), 412.



10.22214/IJRASET

45.98

IMPACT FACTOR: 7.129

IMPACT FACTOR: 7.429

INTERNATIONAL JOURNAL FOR RESEARCH

IN APPLIED SCIENCE & ENGINEERING TECHNOLOGY

Call : 08813907089 (24*7 Support on Whatsapp)