

ECG Signal Processing using Multiple Techniques: A Review

Neha Vikas Khandait¹, Asst. Prof. Mr A.A. Shirolkar²

¹Electronics Technology, Department of Technology, Shivaji University, Kolhapur, Maharashtra, India

²Department of Technology, Shivaji University, Kolhapur, Maharashtra, India

Abstract: *Electrocardiogram (ECG) is one the important biomedical signal. ECG signal consists of different segments as QRS complex, ST segment and PR segment. Features of an ECG signal are nothing but these segments and intervals between fiducially points such as RR interval, amplitude of P, R and T wave. The ECG signal is corrupted by artefacts, this artifacts are produced by different sources of like artificial and biological nature. Main artefacts are power line interference, Impulse noise, Electrostatic potentials and noise of electronic devices which are artificially produces. Motion artefacts and muscle artefacts (EMG signal) are the main biological artefacts. This paper introduces the digital filtering method to cope with the noise artefacts in the ECG signal. Multiple filters are applied on the ECG signal to reduce artefacts from ECG. ECG attracted numerous researchers because of its pertinence for determining the condition of human's heart. Specific measurements using the ECG are commonly used by medical practitioners to portend early symptoms of heart disorders. Still, these measurements are concerned by an unwanted artefact which cannot be eliminated using simple filtering methods. Numerous studies have been conducted to develop ECG de-noising techniques. According to this context, following review is undertaken.*

Keywords: *ECG, PQRST wave, QRS complex, artefact's removal*

I. INTRODUCTION

Electrocardiography (ECG) is the mechanism of recording the electrical activity of the heart for the period using electrodes, which are placed on the skin. the tiny electrical changes on the skin of body that emerge from the heart muscles are detected by Those electrodes. The pattern produces during depolarizing and repolarizing during a heartbeat is called Electro physiologic. It is general performed to detect any cardiac problems. A typical ECG recording consists of P-wave, T-wave, U-wave and QRS-complex. The main combination of three waves is seen on a typical electrocardiogram its name is QRS complex. The central and most visually conspicuous part of the tracing is the QRS complex; in other words, it's the main spike seen on an ECG line. The above all problems are related to the depolarization of the right and left ventricles of the human heart and contraction of the large ventricular muscles. To measure various cardiac conditions The amplitude and interval of separate waves are used. The most intrinsic interval used for cardiovascular abnormality stipulation is RR-interval (distance between two R-peaks). an normal RR-interval is 600-1200 ms. an normal ST-segment and T-wave duration is 80-120 ms and 120-160 ms respectively which are useful for heart diseases. The heart rate (HR) is computed from the extracted features of ECG signal. The heart rate can be useful to identify various cardiac abnormalities like Bradycardia, Tachycardia, Bundle Branch Block, Premature Ventricular Contraction, Wolff-Parkinson-White syndrome (WPW), etc.,

The depravity detection process takes more time if figure out by doctor using long span ECG data. So, an automatic technique is necessary for early detection of cardiovascular diseases. The ECG beat classification is essential for automatic detection and diagnosis of cardiovascular diseases in a long duration electrocardiogram.

Electrocardiogram (ECG) indicates electrical activity of human heart. ECG consists of 5 waves - P, Q, R, S and T. This signal could be calculated by electrodes from human body in ideal engagement. Information from these electrodes are brought to simple electrical circuits with amplifiers and analogue digital promoter. power supply network 50 Hz frequency and breathing muscle are the artefacts are interface with the ECG signal this is the main problem of digitalized signal is interference with other noisy signals. These noisy signals have to be eliminated before the signal is used for next data processing like heart rate frequency unmasking. For real-time applications in embedded devices the Digital filters and signal processing should be composite. Heart rate frequency is very essential health status information. The frequency assessment is used in many medical or sport applications like stress tests or life treating condition prediction. It is the possible way to get heart rate frequency is computed it from the ECG signal. Heart rate frequency can be measure from ECG signal using many methods and algorithms. Multiple algorithms for heart rate detection are based on QRS complex detection and heart rate is figure out like distance between QRS complexes. QRS complex can be measure

using for example algorithms from the field of simulated neural networks, genetic algorithms, wavelet transforms or filters banks. The direct way how to find QRS complex is to use an adaptive threshold. The straightforward methods for heart rate detection are ECG signal spectral analyse and Short-Term Autocorrelation method.

II. LITERATURE SURVEY

One of the previous systems proposed by Udit Satija, Barathram. Ramkumar, and M. Sabarimalai Manikandan [1] was designed the method modified ensemble empirical mode decomposition (CEEMD), the short term temporal feature extraction, and the decision rules based noise detection and classification. In this method, ECG signals are first decomposed using the modified CEEMD algorithm for discriminating the ECG components from the noises and artefacts. Then, the short-term temporal features such as maximum absolute amplitude, number of zero crossings, and local maximum peak amplitude of the autocorrelation function are computed from the extracted high-frequency (HF) and low frequency (LF) signals. Finally, a decision rule-based algorithm is presented for detecting the presence of noises and classifying the processed ECG signals into six signal groups: noise-free ECG, ECG+BW, ECG+MA, ECG+PLI, ECG+BW+PLI, and ECG+BW+MA. The proposed framework achieves better noise detection and classification rates than the current state-of-the-art methods along with accurately localizes short bursts of noises with low endpoint delineation errors

An artifact cancellation with adaptive neuro-fuzzy inference system (ANFIS) is introduced for denoising ECG signal was developed by Salem Belgurzi, Ibrahim Elshafiey and Adnan Nouh [2] in this ECG signaled detection has been commonly used by individuals and doctors inside and outside the clinic to diagnose cardiac diseases. For the correct diagnosis, artifacts removal that they affect the quality of ECG signal becomes needful. In this paper, the ANFIS method is firstly applied to suppress the power line interference and the baseline wander artifacts from corrupted ECG signal then the outputs of the proposed method are compared with the outputs of methods such as LMS (the least mean Square) and NLMS (Normalized the Least Mean Square). Based on their experiments that were performed on the number of ECG signals, it was shown that the proposed ANFIS method can significantly detect and remove the artifacts In this paper, an algorithm based on an adaptive neuro-fuzzy inference system (ANFIS) is proposed to extract the ECG signal from a contaminated ECG signal. They focus on denoising two artifacts that contaminate the ECG signal; namely power line interference, and baseline wandering. the performance of the ANFIS method, the SNR improvements were calculated. The SNR involved in this approach was analyzed and compared with other methods. The comparison results showed that the ANFIS method was successful for the estimation of the corrupted ECG signal by artifacts.

An Extended Kalman Filter (EKF) has been proposed for the filtering of noisy ECG signals was developed by R Sameni¹, MB Shamsollahi, C Jutten, M Babaie-Zadeh [3] This method was depend on a modified nonlinear dynamic model, prior introduced for the generation of synthetic ECG signals. An automatic parameter selection method has also been recommended, to adapt the model with a vast variety of normal and abnormal ECG signals. The outcomes show that the EKF output is able to find the original ECG signal shape even in the noisiest epochs of the ECG signal. The EKF method may serve as an efficient filtering process for applications such as the noninvasive extraction of fetal cardiac signals from maternal abdominal signals. The EKF's dynamic model was depend on a modified three dimensional nonlinear dynamic model prior introduced for the generation of synthetic ECG signals. This nonlinear model was linearized in order to be used in an proposed method. The EKF filter was after applied to noisy ECG signals, and the outcomes show the filter's capability in tracking and filtering noisy ECG signals.

Wavelet transform is method that developed for point detection from ECG signal by S. ktata, k. Ouni and N.Ellouze [4] in this detect automatically the R peaks, the T and P wave maxima, separately. After having represented the ECG equivalent in time frequency domain, they detect the complex QRS maximum and the T wave using the truncation of these waves by rectangular window. The influence of scale levels in their algorithm is tested for a large amount of ECG signals. Firstly, the original ECG signal is decomposed on five scale based on Daubechies wavelet. The best detection of the complex QRS is given for the high frequencies equal to level 1 because of the short time transition. Wile, the ECG characteristics are detected a rectangular window is applied in order to truncate every single QRS and T wave. Finally, they concatenate the detected features maxima obtained with their approach on the original signal. This adopted methodology gives more accuracy where it's employed on the transformed ECG signal one. The detection of the P and T waves, QRS complex in original ECG is a difficult problem due the time varying morphology and the non-stationary of physiological signals.

Projection operator based approach to remove baseline wander noise from the ECG signal was evolved by Sakshi Agrawal and Anubha Gupta [5] The artefact's subspace is generated using sample functions of the first order fBm procedure characterizing baseline wander artefact's. The orthogonal projection of corrupted ECG signal onto the noise subspace provides an estimate of baseline wander noise. The calculated noise is removed from the noisy input signal to obtain noise-free ECG signal. Performance

analyse with other baseline wander removal technics shows the efficacy of the proposed method. Real-time ECG signals are usually corrupted by baseline wander artefact's which needs to be eliminate before an ECG signal can be used by a doctor for analysis. In proposed method removed baseline wander noise from noisy ECG signals. They used the fBm modeling of baseline wander noise to create a projection operator based approach to remove it from noisy ECG signals. The result of these techniques is output signal-to-noise ratio at poor input SNR and also leads to higher value of cross correlation coefficient between the original (clean) ECG and the noise free ECG signal.

Rebeh mabrouki, Balkine khaddoumi and Mounir sayadi [6] present a combination between Empirical Mode Decomposition (EMD) approach and Hilbert transform approach for the purpose of R peak detection in Electrocardiogram (ECG) signal. This technic uses the EMD to find out the signal which highlights the region of the QRS complex in ECG signal by associates the first three IMF that consist of sufficient information about the region of the QRS complex then the envelope obtained from Hilbert transform to detect the R peaks. This method requires the some stages: removed the baseline wander from the original ECG signal, separate the resulting filtered ECG signal into a collection of AM-FM components called Intrinsic Mode Functions (IMF) which are achieve by using Empirical Mode Decomposition approach, sum the first three Intrinsic Functions Mode (IMFs) which accommodate enough information about the QRS complex, determine the first derivative of the sum signal to get the points of minima or maxima, The differentiated signal is then transformed using Hilbert transform and then calculate the Hilbert envelope, and finally, find the positions of the maximum which indicates the positions of the R peaks.

Chavdar Levkov, Georgy Mihov, Ratcho Ivanov, Ivan Daskalov, Ivaylo Christov and Ivan Dotsinsky [7] was developed the Subtraction procedure which removed power-line interference from the ECG signal without affecting its spectrum. The metod operates successfully with amplitude and frequency deviations of the obstruction. The frequency deviations are first alone by hardware measurement of the power-line frequency. Software analyse of the interference period was improve for battery supplied units and some ECG modules connected to personal computers. The developed structure of on-line going subtraction procedure leads to its extended implementation careless of multiplicity between sampling rate and interference frequency. The structure is adaptive due to the introduced filtering modules, are called into use depending on the type of sampling. The presented scrutinizing of the elimination procedure and the different types of notch filters approve the advantages of this method for interference cancellation in ECG signals.

Omid Sayadi 1, Mohammad Bagher Shamsollahi 2 [8] are presented a new method for ECG baseline correction using the adaptive bionic wavelet transform (BWT). Actually by the means of BWT, the resolution in the time-frequency domain can be adaptively adjusted not only by the signal frequency but also by the signal instantaneous amplitude and its first-order differential. Besides by optimizing the BWT parameters parallel to modifying previous thresholding rule, one can handle ECG baseline correction. First calculation of the baseline wandering frequency is obtained and then the adaptation can be used only in three successive scales in which the mid-scale has the closest centre frequency to the calculated frequency. Hence the implementation is time consuming.

Yue-Der Lin and Yu Hen Hu [9] developed power-line interference (PLI) detection method and suppression algorithm is presented to pre-process the electrocardiogram (ECG) signals. A distinct feature of this introduced algorithm is has capability to detect the presence of PLI in the ECG signal before applying the PLI suppression algorithm. No PLI suppression operation will be done if PLI is not detected. They suggested a PLI detector that employs an optimal linear discriminant analysis (LDA) algorithm to take a stands for the PLI presence. An efficient recursive least-square (RLS) adaptive notch filter is also evolved to serve the purpose of PLI suppression. In this PLI detection and suppression algorithm is easily implemented and computationally efficient. Results shows that the PLI algorithm can effectively detect and suppress the presence of PLI in ECG signals without human operator supervision, even in the case of variant power-line frequency. The potential applications of this algorithm are not limited to ECG signals only and may have profound impact on numerous biomedical instruments.

Hyejung Kim, Refet Firat Yazicioglu, Patrick Merken, Chris Van Hoof, and Hoi-Jun Yoo [10] was develop an ECG signal processing method with quad level vector (QLV) for the ECG holter system. The ECG processing made up of of the compression flow and the classification flow, and the QLV is suitable for both flows to gain better achievement with low-computation complications. The compression algorithm is execute by using ECG skeleton and the Huffman coding. To reduce the processing cost while maintaining the signal quality some methods are used like Unit block size optimization, adaptive threshold adjustment, and 4-bit-wise Huffman coding. The heartbeat segmentation and the R-peak detection technics are employed for the classification algorithm and the noise robust test is also execute for the reliability of the algorithm. Its average compression ratio is 16.9:1 with 0.641% percentage (RMS) root mean square difference value and the encoding rate is 6.4 kbps. The efficiency performance of the R-peak detection is 100% without noise and 95.63% at the worst case with -10-dB SNR noise. The overall processing cost is minimize by 45.3% with the compression techniques.

III. CONCLUSION

After reviewing all the above research articles, it can be concluded that, the noise removing from ECG signal is essential in biomedical. The main use of ECG is to detect the heart disease by using signals generated by heart. ECG signal contains different noise signals which are removed by using different algorithms mentions in above. This paper step towards the removing noise from ECG signal and after that also indicates the deceases

REFERENCES

- [1] Udit Satija, Barathram.Ramkumar, and M. Sabarimalai Manikandan; Automated ECG Noise Detection and Classification System for Unsupervised Healthcare Monitoring; 2016 IEEE Journal of Biomedical and Health Informatics
- [2] Salem Belgurzi, Ibrahim Elshafiey and Adnan Nouh; Artifacts Removal from ECG signal using an ANFIS Technique
- [3] R Sameni, MB Shamsollahi, C Jutten, M Babaie-Zadeh; Filtering Noisy ECG Signals Using the Extended Kalman Filter Based on a Modified Dynamic ECG Model, 2005 IEEE, 32:1017–1020
- [4] S. ktata, k. Ouni and N.Ellouze; ECG Signal Maxima Detection Using Wavelet Transform, 1-4244-0497-5/06 C 2006 IEEE
- [5] Sakshi Agrawal and Anubha Gupta; Projection Operator Based Removal of Baseline Wander Noise from ECG Signals; 978-1-4799-2390-8/13 203 IEEE
- [6] Rebeh MABROUKI, Balkine KHADDOUMI and Mounir SAYADI; R Peak Detection in Electrocardiogram Signal Based on a Combination between Empirical Mode Decomposition and Hilbert Transform; 1st International Conference on Advanced Technologies for Signal and Image Processing - ATSIP2014 March 17-19, 2014,
- [7] Chavdar Levkov¹, Georgy Mihov², Ratcho Ivanov², Ivan Daskalov³, Ivaylo Christov³ and Ivan Dotsinsky^{*3}; Removal of power-line interference from the ECG: a review of the subtraction procedure; BioMedical Engineering OnLine 2005, online.com/content/4/1/50 <http://www.biomedical-engineering>
- [8] Omid Sayadi 1, Mohammad Bagher Shamsollahi 2; ECG Baseline Correction with Adaptive Bionic Wavelet Transform; 1-4244-0779-6/07 2007 IEEE
- [9] Yue-Der Lin and Yu Hen Hu*; Power-Line Interference Detection and Suppression in ECG Signal Processing; IEEE transactions on biomedical engineering, vol. 55, no. 1, january 2008
- [10] Hyejung Kim, Refet Firat Yazicioglu, Patrick Merken, Chris Van Hoof, and Hoi-Jun Yoo; ECG Signal Compression and Classification Algorithm With Quad Level Vector for ECG Holter System; IEEE transaction on information technology in biomedicine, vol. 14, no. 1, January 2010