



iJRASET

International Journal For Research in
Applied Science and Engineering Technology



INTERNATIONAL JOURNAL FOR RESEARCH

IN APPLIED SCIENCE & ENGINEERING TECHNOLOGY

Volume: 6

Issue: XI

Month of publication: November 2018

DOI:

www.ijraset.com

Call:  08813907089

E-mail ID: ijraset@gmail.com

Atomic Absorption Spectrophotometric Analysis of Lead in Water

Shobharani Panchagnula

Department of chemistry, G.Narayanamma Institute of Technology and Science (Autonomous), Telangana, India.

Abstract: *The aim of the present work was to test the application of atomic absorption spectroscopy, to the analysis of lead in various water samples.*

Traces of lead found in water can be analysed by atomic absorption spectrophotometer by extraction with dithiocarbamate followed by subjecting to the flame of atomic absorption spectrophotometer. The above mentioned method avoids the inaccuracies of many conventional methods. In the present paper an attempt was made to focus mainly on the determination of lead in water samples. Atomic absorption spectrophotometer entails the determination of the absorption of the metal at its line centre by using a narrow line source emitting the given resonance line whose emission line profile is less than the absorption line profile of the analyte in its flame.

Earlier methods involved direct extraction with a complexing agent or colour development with suitable reagents under controlled conditions of pH. which were time consuming involving other difficulties. Results of investigations were accurate and sensitive and free from interferences by extraneous elements.

Keywords: *atomic absorption spectrophotometer, dithiocarbamate*

I. INTRODUCTION

In the recent decades pollution from heavy metals has become a matter of serious concern. Arsenic mercury, copper cadmium and lead are some of the trace elements that cause highest risk to human health(1).

High levels of lead contamination in the blood in children may have a significant impact in lowering their IQ and increasing the risk of other diseases. Sources of high levels of lead in the blood are lead based paints, lead contaminated dust, water, air and soil. It is also contracted from the adults who work with batteries, work in home renovations, automobile repair shops and processed foods(2). Development and learning delays are some of the early symptoms of contamination along with slowed growth and mood disorders(3). Apart from atmospheric lead, the use of lead in industries which makes its way through water into the human system was found to be high. Gastrointestinal problems like vomiting, constipation and severe fatigue are seen. Chronic low lead exposure has also been associated with cognitive impairment, cardiovascular disease, anemia, low birth weight and premature mortality. The implications of exposure are extremely detrimental (4,5). Lead enters into the atmosphere through industrial waste, human activities, mining, vehicle emissions, lead acid batteries, fertilizers, paints etc. and thereby enter human tissues through diet. Heavy metals are hard to metabolise and bioaccumulate in the bodies which causes many health hazards(6).

II. METHODOLOGY

The metal atoms absorb strongly at characteristic wavelength which coincide with the emission spectral lines of the particular metal. Atomic absorption spectrophotometry entails the determination of the element at its line centre by using a narrow line source emitting the given resonance line whose emission line profile is less than absorption line profile of its analyte in flame. The free atoms generated in the atomiser absorb radiation at a specific frequency (7,8). The flame gases are treated as a medium containing free unexcited atoms capable of absorbing radiation from an external source when the radiation corresponds exactly equal to the energy required for transition of the test element from the ground electronic state to an upper excited electronic state. Unabsorbed radiation passes through a monochromator that isolates the excited spectral line. Absorption is measured by the difference in transmitted signal in presence of absorbance of the test element.

Atomic absorption spectrometer uses the absorption of light to measure the concentration of gas phase atoms. The light that is focussed into the flame is produced by a hollow cathode lamp, inside which is the sample. The samples were vaporised in a flame of graphite furnace. Standards of the heavy metal were prepared and vaporised. Standard solutions were made by dilution containing 100mg per litre of the heavy metal, by dissolving over dried lead nitrate in the minimum quantity of hydrochloric acid. Working standards were made up from these as necessary.

Traces of the absorbed heavy metals in the atomiser and spray chamber were removed before each measurement by spraying nitric acid and water. A fine wire was passed through the atomiser capillary after every 5 measurements to clear the deposited matter. Mercury hollow cathode lamp was used as the light source. The light emitted by the cathode was focussed at the centre of the flame into which the sample was aspirated and focussed into the monochromator set to pass the resonance line. The signal from the photomultiplier behind the monochromator was amplified by an amplifier tuned at 50 cycles per second and the rectified output was read on a microammeter (9,10).


III. RESULTS AND DISCUSSION

A compressor and a cylinder fitted with acetylene gas is used. The pressure is set to 1 psi. Hollow cathode lamp of the sample element is used. A calibration graph is obtained by plotting absorbance Vs. concentration as shown in figure-1 and the data is as shown in table-1. There was ease of sample pre treatment compared to other methods. There was improved sensitivity and an increase in detection power and an increase in reproducibility of the results.

TABLE-1

S.No.	Volume of stock	Concentration (ppm)	Volume of double distilled water	Absorbance
1	0.5	10	49.5	0.2
2	1.0	20	49	0.32
3	1.5	40	48.5	0.68
4	2.0	60	48.0	0.83
5	2.5	80	47.5	0.97

FIGURE-1



IV. CONCLUSION

Earlier methods involved direct extraction with a complexing agent or colour development with suitable reagents under controlled conditions of pH (11,12). Matrix interferences and high background absorption have been reduced in this method. Those methods were time consuming involving other difficulties. Results of investigations were accurate and sensitive and free from interferences by extraneous elements (13). There was greater reproducibility and was found to be relatively less expensive approach and the sample consumption was also low when compared to other methods (14,15).

V. ACKNOWLEDGMENTS

The author is thankful to the Chairman, Joint secretary and Principal of G.Narayanamma Institute of Technology and Science for providing all the facilities to carry on the research work.

REFERENCES

- [1] Determination of lead in blood using electrothermal atomisation atomic absorption spectrometry with a l'vov platform and matrix modifier. Dayton T. Miller, Daniel C. Paschal, Elaine W. Gunter, Phillip E. Stroud and Joseph Analyst issue 4.
- [2] Determination of lead and other heavy metals in urine by Atomic Absorption Spectroscopy. J. B. Willis . Anal. Chem., 1962, 34 (6), pp 614–617.
- [3] Determination of accurate extinction coefficients and simultaneous equations for assaying chlorophylls a and b extracted with four different solvents: verification of the concentration of chlorophyll standards by atomic absorption spectroscopy. IR.J. Porra W.A. Thompson*P.E. Kriedemann* Biochimica et biophysica acta (BBA)- Bioenergies. Volume 975, August 1989, Pages 384-394.
- [4] Rapid determination of lead in water samples by dispersive liquid–liquid microextraction coupled with electrothermal atomic absorption spectrometry. MohammadTaghiNaseriabMohammad RezaMilani Hosseiniab Yaghoub Assadiab Armin Kianiab Talanta Volume75, issue 1, March 2008, Pages 56-62.
- [5] Cloud point extraction and flame atomic absorption spectrometric determination of cadmium(II), lead(II), palladium(II) and silver(I) in environmental samples.
- [6] Flow injection on-line separation and preconcentration for electrothermal atomic absorption spectrometry. Part 1. Determination of ultratrace amounts of cadmium, copper, lead and nickel in water samples. Michael Sperling, Xuefeng Yin and Bernhard Welz. Journal of Hazardous Materials. Volume 168, Issues 2–3, 15 September 2009, Pages 1022-1027.
- [7] The determination of some heavy metals in food samples by flame atomic absorption spectrometry after their separation-preconcentration on bis salicyl aldehyde, 1,3 propan diimine (BSPDI) loaded on activated carbon M.GhaediaA.ShokrollahiaA.H.KianfaraA.S.MirsadeghiaA.PourfarokhiaM.Soylakb.
- [8] Cloud point extraction for the determination of cadmium and lead in biological samples by graphite furnace atomic absorption spectrometry. Tatiane de A.Maranhão Daniel L.G. Borgesa Márcia A.M.S.da Veigab Adilson J.Curtiusa. Spectrochimica Acta Part B: Atomic Spectroscopy. Volume 60, Issue 5, 30 June 2005, Pages 667-672
- [9] Determination of lead and nickel in environmental samples by flame atomic absorption spectrometry after column solid-phase extraction on Ambersorb-572 with EDTA Journal of Hazardous Materials. Volume 129, Issues 1–3, 28 February 2006, Pages 130-136.SitkiBaytakaA. Rehber Türkerb.
- [10] On-line separation and preconcentration of lead(II) by solid-phase extraction using activated carbon loaded with xylenol orange and its determination by flame atomic absorption spectrometry. Ali A.EnsafiA. ZendegiShiraz. Journal of Hazardous Materials Volume 150, Issue 3, 11 February 2008, Pages 554-559.
- [11] Determination of As, Cd, Cu, Hg and Pb in biological samples by modern electrothermal atomic absorption spectrometry. JordiSardansa Fernando Montesb Josep Peñuelasa Spectrochimica Acta Part B: Atomic Spectroscopy Volume 65, Issue 2, February 2010, Pages 97-112.
- [12] An efficient flow-injection system with on-line ion-exchange preconcentration for the determination of trace amounts of heavy metals by atomic absorption spectrometry. ZhaolunFangIJRůžičkaE.HHansen. Analytica Chimica Acta, Volume 164, 1984, Pages 23-39.
- [13] Determination of lead in gasoline by atomic absorption spectroscopy. L Masayuki Kashiki Seigo Yamazoe Shozo Oshim Analytica Chimica Acta Volume 53, Issue 1, January 1971, Pages 95-100.
- [14] Direct determination of Cu, Mn, Pb, and Zn in beer by thermospray flame furnace atomic absorption spectrometry☆
- [15] Clésia C. Nascentes Marcos Y. Kamogawaab Kelly G. FernandesMarco A.Z. Arrudac Ana Rita A. Nogueirab Joaquim A. Nóbregaa Spectrochimica Acta Part B: Atomic Spectroscopy Volume 60, Issue 5, 30 June 2005, Pages 749-75.
- [16] Study on stable coatings for determination of lead by flow-injection hydride generation and in situ concentration in graphite furnace atomic absorption spectrometry☆ H.O. Haug, Spectrochimica Acta Part B: Atomic Spectroscopy. Volume 51, Issue 11, September 1996, Pages 1425-1433



10.22214/IJRASET



45.98



IMPACT FACTOR:
7.129



IMPACT FACTOR:
7.429



INTERNATIONAL JOURNAL FOR RESEARCH

IN APPLIED SCIENCE & ENGINEERING TECHNOLOGY

Call : 08813907089  (24*7 Support on Whatsapp)