



# **iJRASET**

International Journal For Research in  
Applied Science and Engineering Technology



---

# **INTERNATIONAL JOURNAL FOR RESEARCH**

IN APPLIED SCIENCE & ENGINEERING TECHNOLOGY

---

**Volume: 7      Issue: VIII      Month of publication: August 2019**

**DOI: <http://doi.org/10.22214/ijraset.2019.8132>**

**[www.ijraset.com](http://www.ijraset.com)**

**Call: ☎ 08813907089**

**E-mail ID: [ijraset@gmail.com](mailto:ijraset@gmail.com)**

# Smart Saline Level Monitoring and Control System

Pooja Pandit Landge<sup>1</sup>, Siddharudha S. Shirgan<sup>2</sup>

<sup>1, 2</sup>N.B. Navale Sinhgad College of Engineering, Solapur,

**Abstract:** *In hospitals, Saline is fed to patients to treat dehydration and thus improve their health. In current health care measures, whenever a saline is fed to any patient, the patient needs to be continuously monitored by a nurse or any caretaker. Almost in all of the hospital, a nurse or caretaker is responsible for monitoring the saline level continuously without any interruptions. Due to the negligence and inattentiveness towards saline completion by doctors, nurses or caretaker of the patients and lack of nurses with sufficient skills in hospitals and their excessive workload, a huge number of patients are being harmed in the hospitals. Hence to prevent the patient from getting harmed and protect their lives during saline feeding period, the saline level monitoring and control system have been developed. The proposed system comprises of sensors which will act as a level sensor for monitoring the critical level of the saline in the saline bottle and control infusion drop rate from using motor mechanism to increasing and decreasing the saline drop rate. The system will show saline droplet status, saline drop rate and remaining time through app to be develop for convenient to hospital staff member. This proposed system can be utilized efficiently in homes as well as hospitals.*

## I. INTRODUCTION

These days technology is growing at a very high speed. Human lives have become much more dependent on electronic devices. Today's world requires sophisticated control in its different electronic gadgets. The basic aim of saline level indicator is to ease human lives.

Automation of the surrounding environment of a modern human being helps to increase the work efficiency and saves time. Saline is fed when the patient's body is dehydrated. A constant monitoring of the saline level in the bottle is required. If the empty saline bottle is not replaced immediately then the pressure difference between the patient's blood pressure and the empty saline bottle causes reverse flow of blood into the saline.

This situation can be a serious threat to the patient's well-being. Thus the automation device is suggested in order to avoid any inconvenience that may be caused to the patient in case of lacking of constant monitoring by patient's relatives or hospital employees.

The nurses can check saline droplet status of each patient's bed without walking around patient's room every hour. The device will send the status to the host unit and show the results. Therefore, nurses can check a line droplet status.

## II. PROBLEM STATEMENT

Almost in all of the hospital, a nurse or caretaker is responsible for monitoring the saline level continuously without any interruptions. Due to the negligence and inattentiveness towards saline completion by doctors, nurses or caretaker of the patients and lack of nurses with sufficient skills in hospitals and their excessive workload, a huge number of patients are dying and are being harmed in the hospitals. Hence to prevent the patient from getting harmed and protect their lives during saline feeding period, we are going to develop the smart saline level monitoring and control system.

## III. OBJECTIVES

- A. To detect the critical level of saline bottle using IR sensor.
- B. To make the system that automatically stop the flow after emptying of saline bottle using motor mechanism to increasing and decreasing drop rate which is controlling through app.
- C. To design android application to display the result in the form of saline droplet rate, remaining time to empty the saline bottle displayed on mobile phone.
- D. Provide cost effective and automatic saline level monitoring and controlling system which can be effortlessly implemented in any hospitals.

#### IV. PROPOSEDSYSTEM




Fig. 1. System Architecture

This proposed system will function for different scenarios which are explained below as follows:

When the system is power on, IR sensor (TX and RX) will detect the drops of fluid and on another aspect, the microcontroller will calculate drop rate (drops per minute). It will also count a number of drops then according to the number of drops; it will calculate the level of fluid and remaining time.




Fig. 2. Position of IR Sensor

An empty infusion solution bag at particular critical set level is achieved then it will alert with notification along with buzzer will be played.

Motor acts as an output for the microcontroller which is a mechanism used for control infusion drop rate.




Fig. 3. Mechanism for stop reverse flow of blood



Fig. 5. System Hardware

## V. RESULT

### A. Android Application

Successfully developed an android application named "Saline Monitoring System" to display the result in the form of saline droplet rate, remaining time to empty the saline bottle displayed on mobilephone.




Fig. 4. Simulation

8:32 AM 2.29K/s 4G VoLTE

### Saline Monitoring System

Login Page

USER NAME :

PASSWORD :

Login

WELCOME

EXIT

Fig. 6. Login Screen of Android Application

8:24 PM 58.1K/s 4G VoLTE

### Saline Monitoring System

#### Saline 1

Saline Condition : Above\_80%

Remaining Time : 40\_to\_45\_min

Drop Rate (DPM) : 15

Moderate Speed

Low Speed

#### Saline 2

Saline Condition : Above\_80%

Remaining Time : 40\_to\_45\_min

Drop Rate (DPM) : 68

Moderate Speed

Low Speed

Fig.7. Result displayed



## VI. CONCLUSION

As the entire proposed system is automated, it requires very less human intervention. It will be advantageous at night. It can wirelessly send the information and display the results in the form of saline droplet rate, remaining time to empty the saline bottle. This will reduce the stress in continual monitoring by the doctor or nurse.

## VII. ACKNOWLEDGMENT

I would like to thank my guide Prof. S. S. Shirgan for his consistent guidance, inspiration and sympathetic attitude throughout the total work, which I am sure, will go a long way in my life. I am grateful for the many useful comments and suggestions provided by him, which have resulted significant improvements in the paper.

## REFERENCE

- [1] Shyama Yadav, Preet Jain Real time cost effective e-saline monitoring and control system International Conference on Control, Computing, Communication and Materials (ICCCCM) February 2016 IEEE 2016.
- [2] Pattarakamon Rangsee, Paweena Suebsombut, Phakphoom Boonyanant Low-Cost Saline Droplet Measurement System using for Common Patient Room in Rural Public Hospital Joint International Conference on Information and Communication Technology, Electronic and Electrical Engineering (JICTEE) January 2014 IEEE 2014.
- [3] H. Amano, H. Ogawa, H. Maki, S. Tsukamoto, Y. Yonezawa and W. M. Caldwell, "A remote drip infusion monitoring system employing Bluetooth," 2012 Annual International Conference of the IEEE Engineering in Medicine and Biology Society, San Diego, CA, 2012, pp. 2029-2032.
- [4] C. F. Huang and J. H. Lin, "A warning system based on the RFID technology for running-out of injection fluid," 2011 Annual International Conference of the IEEE Engineering in Medicine and Biology Society, Boston, MA, 2011, pp. 2212-2215.
- [5] Priyadarshini. R, Mithuna. S, Vasanth Kumar. U, Kalpana Devi. S, Dr. Suthanthira Vanitha. N Automatic Intravenous Fluid Level Indication System for Hospitals International Journal for Research in Applied Science & Engineering Technology (IJRASET) Volume 3 Issue VIII, August 2015 Pg.no : 427-432.
- [6] P. Pearlina Sheeba, N. Anushree, L. Aishwarya Saline Infusion Level Detection and Heart Rate Monitoring System International Journal for Research in Applied Science & Engineering Technology (IJRASET) Volume 4 Issue XI, November 2016 Pg.no: 637-641.
- [7] Manoj Kumar Swain, Santosh Kumar Mallick, Rati Ranjan Sabat Smart Saline Level Indicator cum Controller International Journal of Application or Innovation in Engineering & Management (IJAIEM) Volume 4, Issue 3, March 2015 Pg.no: 299-301.
- [8] C.C. Gavimath, Krishnamurthy Bhat, C.L. Chayalakshmi, R.S. Hooli, B.E. Ravishankera Design and development of versatile saline flow rate measuring device and GSM based remote monitoring device International Journal of Pharmaceutical Applications (IIPA) Volume 3, Issue 1, 2012, Pg.no: 277-281.
- [9] P. Kalaivani, T. Thamaraiselvi, P. Sindhuja and G. Vegha Real Time ECG and Saline Level Monitoring System Using Arduino UNO Processor Asian Journal of Applied Science and Technology (AJAST) Volume 1, Issue 2, March 2017 Pg.no: 160-164.
- [10] Mansi G. Chidgopkar, Aruna P. Phatale Automatic and low cost saline level monitoring system using wireless bluetooth module and CC2500 transceiver International Journal of Research in Engineering and Technology (IJRET) Volume: 04 Issue: 09 September-2015 Pg.no: 274-276.
- [11] I.S. Tawade, M.S. Pendse, H.P. Chaudhari Design and Development of Saline Flow Rate Monitoring System Using Flow Sensor, Microcontroller and RF ZigBee Module International Journal of Engineering Research and General Science (IJERGS) Volume 3, Issue 3, May-June, 2015 Pg.no: 472-478.
- [12] R. Vasuki, Dennis and Hem Priya Chander Designing a portable monitoring device to measure the drips rate International Journal of Biotechnology Trends and Technology (IJBT) Volume 1 Issue 3 Nov-Dec 2011 Pg.no: 29-35.



10.22214/IJRASET



45.98



IMPACT FACTOR:  
7.129



IMPACT FACTOR:  
7.429



# INTERNATIONAL JOURNAL FOR RESEARCH

IN APPLIED SCIENCE & ENGINEERING TECHNOLOGY

Call : 08813907089  (24\*7 Support on Whatsapp)